

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMi Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: A.J. Nederlof

### Applicant

Name: Mettler-Toledo, Inc.  
Address: 1150 Dearborn Drive  
Worthington OH 43085-6712  
United States of America

### Manufacturer of the certified pattern

Name: Mettler-Toledo, Inc.  
Address: 1150 Dearborn Drive  
Worthington OH 43085-6712  
United States of America

### Identification of the certified pattern

Type : 0743  
Class : C  
 $P_i = 0.7$

Maximum Capacity ( $E_{max}$ )	9072 kg, 13608 kg and 20412 kg 20000 lb, 30000 lb and 45000 lb	
Accuracy Class	C	
Maximum number of LC intervals (n)	2000	3000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	6000	



This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC2977 and the appertaining documentation folder, with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60  
edition 1991 (E)  
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.


The conformity was established by tests described in the associated test report  
N° R60/1991-NL-97.12, that includes 37 pages;  
N° R60/1991-NL-97.31, that includes 37 pages.

The issuing authority  
A.J. Nederlof  
*A.P.*



1 December 1997

The OIML member  
G.J. Faber



1 December 1997

\*  
\* \*