



Certificate of Compliance

Certificate: 1413503 (LR 55000C)

Master Contract: 173467

Project: 1596158

Date Issued: 2004/11/10

Issued to: Mettler-Toledo GmbH

Im Langacher
Greifensee, 8606
Switzerland
Attention: Mr. A. Atland

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Eknath Desai; C.E.T.

Authorized by: Gianluca Arcari, P.Eng.,
Product Group Manager

PRODUCTS

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

CLASS 8721 04 - LABORATORY EQUIPMENT - Electrical

Clean and Calibration system, Model Easy Clean 100, Easy Clean 150, Easy Clean 200, Easy Clean 300, Easy Clean 350; rated input: 100-230V; 50/60Hz; 0.18-0.3A, output: 24V dc, 1.3A; Class I; pollution degree 2, installation category III, fixed mobility, permanently connected, continuous operation, IP65.

APPLICABLE REQUIREMENTS

CAN / CSA - C22.2 No 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control,

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.