

Member State of OIML  
Germany



OIML Certificate N°  
R76/1992-DE-03.03  
Revision 1

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, D-38116 Braunschweig  
Person responsible: Dr. Roman Schwartz

### Applicant:

Name: Mettler-Toledo GmbH  
Address: Im Langacher, 8606 Greifensee  
Switzerland

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern: Non-automatic electromechanical weighing instrument  
type **PL...-S**

Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 1992, including Amendment 1 (1994),  
for accuracy class **II**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

OIML Certificate N°  
R76/1992-DE-03.03  
Revision 1

The conformity was established by tests described in the associated test report N° 1.14 - 03000215

The issuing authority



Dr. R. Schwartz  
Direktor und Professor

2003-10-24



The OIML member



Prof. Dr. M. Kochsiek  
Vizepräsident

2003-10-24

Identification of the pattern (continued)

Compact weighing instrument with built-in display, keyboard and data-interface.

The weighing ranges with Max, Min, e and number of verification scale intervals may be chosen within the limits of No. 3.2 of R 76-1 and of table 1.

Table 1

Accuracy class	II
Max	81 g ... 8100 g
Verification scale interval e =	0,01 g ... 1 g
Number of verification scale intervals n ≤	15500
Temperature range	10 °C ... 30 °C

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.