

Issued by NMI Certin B.V.

In accordance with WELMEC 8.8 Issue 2, Paragraph 8.1 of EN 45501:1992/AC:1993, OIML R60:2000, WELMEC 2.4 Issue 2.

Producer Mettler-Toledo (Changzhou) Precision Instrument Ltd.
No.5, Middle Huashan Road, Xinbei District
Changzhou
China

Measuring instrument A **bending beam load cell**, with strain gauges, tested as a part of a weighing instrument.

Brand : Mettler-Toledo (Changzhou)
Precision Instrument Ltd.

Designation : SLP845

Further properties are described in the annexes:

- Description TC8077 revision 0
- Documentation folder TC8077-1

An overview of performed tests is given in the annex:

- Description TC8077 revision 0

Issuing Authority **NMI Certin B.V.**
20 December 2012

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" www.nmi.nl)

Reproduction of the complete document only is permitted



Description

Number **TC8077** revision 0
Project number 11200876
Page 1 of 3

1 General information about the load cell

All properties of the load cell, whether mentioned or not, shall not be in conflict with the standards mentioned in this certificate.

This certificate is the positive result of the applied voluntary, modular approach, for a component of a measuring instrument, as described in WELMEC guide 8.8. The complete measuring system must be covered by an EC type-examination Certificate.

1.1 Essential parts

Number	Pages	Description	Remark
8077/0-01	1	Electrical and mechanical diagram	-

Cable:

- The load cell is provided with a 6-wire system (=“Remote-sensing”):
 - The cable length is not limited.

The cable should be a shielded cable, the shield is not connected to the load cell.

1.2 Essential characteristics

Maximum capacity (E_{max})	15 kg up to and including 200 kg
Minimum dead load	0,1 kg
Accuracy Class	C
Rated Output	2,0 mV/V
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	16000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	3000
Input impedance	1116 $\Omega \pm 20 \Omega$
Temperature range	-10 °C / +40 °C
Fraction p_{LC}	0,7
Humidity Class	CH
Safe overload	150% of E_{max}
Output impedance	1000 $\Omega \pm 3 \Omega$
Recommended excitation	5 - 15 V DC
Excitation maximum	20 V DC
Transducer material	Stainless steel
Atmospheric protection	Hermetically welded

The characteristics for n_{max} and Y can be reduced separately. Z is proportional or equal to n_{max} .

Each produced load cell is provided with an accompanying document with information about its characteristics.

1.3 Essential shapes

The load cell is built according to drawing:

- Electrical and Mechanical diagram, drawing number 8077/0-01.

The data plate is secured against removal by sealing or will be destroyed when removed. The data plate mentions at least the information and markings as described in the OIML R60 recommendation.

In the countries where it is mandatory the load cell should bear this test certificate number: TC8077.

2 Seals

The connecting cable of the load cell or the junction box is provided with possibility to seal.



Description

Number **TC8077** revision 0
Project number 11200876
Page 3 of 3

3 Conditions for conformity assessment

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in WELMEC 2 Issue 5 Section 11, at the time of EC verification or declaration of EC conformity of type.

Other parties may use this Part Certificate without the written permission of the producer.

4 Test reports, evaluation reports and pattern evaluation reports

An overview of performed tests is given in the reports:

- No. NMI-11200876-01 dated 14 December 2012 that includes 46 pages;
- No. NMI-11200876-02 dated 14 December 2012 that includes 49 pages;
- No. NMI-11200876-03 dated 14 December 2012 that includes 7 pages.