



FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

0713, 0736, 0744, 0745, 0757, 0790. Load Cell.

IS / I, II, III / 1 / ABCDEFG - TA800821; Entity; NI / 1 / 2 / ABCD; S / II, III / 2 / FG
Refer to drawing TA800821 for entity parameters

Equipment Ratings:

Intrinsically Safe for use in Class I, II and III, Division 1, Groups A, B, C, D, E, F and G when installed in accordance with Entity requirements and control drawing TA800821; Nonincendive for use in Class I, Division 2, Groups A, B, C and D; Suitable for use in Class II and III, Division 2, Groups F and G Hazardous (Classified) Locations

Approved for:

Mettler-Toledo Inc.
1900 Polaris Parkway
Columbus, OH 43240



This certifies that the equipment described has been found to comply with the following FM Approval Standards and other documents:

Class 3600	1998
Class 3610	1988
Class 3810	1989

Original Project ID: 3013019

FM Approval Granted: July 8, 2002

Subsequent Revision Reports / Date FM Approval Amended

FM Global Technologies LLC

A handwritten signature in cursive script, reading "David W. Styracula", written over a horizontal line.

David W. Styracula
Technical Team Manager

7/17/02
Date