

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-03.15

Project number 303135

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: P.P.M. van Enkevort

Applicant

Name: Mettler-Toledo, Inc.
Address: 1150 Dearborn Drive
Worthington, Ohio 43085-6712
United States of America

Manufacturer of the certified pattern

Name: Mettler-Toledo, Inc.
Address: 1150 Dearborn Drive
Worthington, Ohio 43085-6712
United States of America

Identification of the certified pattern

Type : PS60

$n \leq 3000$ divisions (per partial weighing range)

$60 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$

$e \geq 20 \text{ g}$

maximum of two partial weighing ranges

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T2879 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

edition 1992

for accuracy class (III)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.


Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-03.15

Project number 303135

Page 2 of 2

The conformity was established by tests described in the associated test reports:

N° R76/1992-NL-96.20, that includes 54 pages.

N° R76/1992-NL-98.16, that includes 21 pages.

N° R76/1992-NL-01.12, that includes 51 pages.

N° R76/1992-NL1-03.15, that includes 19 pages.

The issuing authority
P.P.M. van Enckevort
Manager Certification Delft

9 May 2003

The CIML member
G.J. Faber

9 May 2003

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.