



# EC type-approval certificate

Number **T5379** revision 9  
Project number 804055  
Page 1 of 6

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body Number 0122

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant Mettler-Toledo GmbH  
Im Langacher  
8606 Greifensee  
Switzerland

In respect of A class **I**, **II** or **III**, electronic, single- or multi-interval **non-automatic weighing instrument**.  
Manufacturer : Mettler-Toledo  
Type : AB-S, AB-L, GB-S, PB-S, PB-L, JB-C, JB-L-C, JB-G and JB-L-G.

Characteristics

|                   | AB-S<br>AB-L        | JB-C <sup>1, 2)</sup><br>JB-L-C <sup>1, 2)</sup> | AB-S                          | GB-S / PB-S / PB-L<br>JB-G / JB-L-G | JB-C <sup>1, 2)</sup> | PB-S           |
|-------------------|---------------------|--|-------------------------------|-------------------------------------|-----------------------|----------------|
| Class             | <b>I</b>            |  | <b>I</b>                      | <b>II</b>                           | <b>II</b>             | <b>III</b>     |
| Max               | ≤ 320 g             | ≤ 1600 ct  | ≤ 220 g                       | ≤ 8100g                             | ≤ 2550 ct             | ≤ 8100 g       |
| e                 | ≥ 1 mg              | ≥ 5 mct  | ≥ 1 mg                        | ≥ 10 mg                             | ≥ 0,1 ct              | ≥ 1 g          |
| d                 | e = d or e = 10d    |  | e = d, e = 10d<br>or e = 100d | e = d, e = 10d or<br>Delta range    |                       | e = d          |
| n                 | ≤ 320000            |  | ≤ 220000                      | ≤ 61000                             |                       | ≤ 8100         |
| Temperature range | +12.5 °C / +27.5 °C |  | +10 °C / +30 °C               | +10 °C / +30 °C                     |                       | +5 °C / +40 °C |

<sup>1)</sup> Can have weighing ranges in classes **I** and **II**

<sup>2)</sup> Weight indication in gram and carat as fixed units.

"-L" is without internal adjustment device

In the description number T5379 revision 9 further characteristics are described.

Valid until 30 September 2018



Nederlands Meetinstituut

# EC type-approval certificate

Number **T5379** revision 9  
Project number 804055  
Page 2 of 6

Description and documentation The instrument is described in the description number T5379 revision 9 and documented in the documentation folder T5379-5, appertaining to this EC type-approval certificate.

Remarks This revision replaces the earlier versions, except for its documentation folder.

Dordrecht, 30 September 2008  
NMI Certin B.V.

Ing. C. Oosterman  
Manager Product Certification

## 1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, may not be in conflict with the legislation.

### 1.1 Essential parts

The electronics;

The mechanical assembly with weighing cell.

EMC protection measures:

- Ferrite bead between the power input and the main board;
- The load receptor and the case of the scale are made of metal;
- The measuring cell is located in a separate metal compartment inside the scale base.

### 1.2 Essential characteristics

Power supply:

- 12 V AC, 50 Hz or
- 12 V DC.

### 1.3 Essential shapes

The non-automatic weighing instrument is built according to the drawings:

- AB-S/ AB-L balances, drawing number ME-240863;
- PB-S/ PB-L/ JB-C balances, drawing number ME-240864;
- GB-S/ PB-S/ PB-L/ JB-G balances, drawing number ME-240865.

The data plate is secured against removal by sealing or will be destroyed when removed.

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing Position of verification- and securing stickers, ME-240868.

The securing component has to bear either:

- A mark of the manufacturer laid down in a notified body approved quality system (Annex II of the directive 90/384/EEC), or
- An official mark of a Member State of the EEC, or another party to the EEA agreement.

Inside the cabinet is a calibration lock, located on the backside of the instrument.

### 1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in article 1(2)(a) of the EC Directive (90/384/EEC), if the peripheral equipment is certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body appointed to certify non-automatic weighing instruments according to paragraph I of Annex II of the EC directive on Non-Automatic Weighing Instruments.

A level indicator which shows that the maximum permissible tilt is being exceeded.

A battery pack that switches the instrument off when the voltage is too low to guarantee correct functioning of the instrument.

## 1.5 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second display's and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in article 1(2)(a) of the EC Directive (90/384/EEC) unless the "preliminary observations" in Annex 1 of this directive is satisfied.
- They do not lead to an instrument having other essential characteristics than those fixed by this type-approval document.

AC/AC-adapter.

## 2 Information about the main constituent parts of the non-automatic weighing instrument

### 2.1 The electronics

#### 2.1.1 Essential parts

| Description   | Drawing number            | Rev.   | Remarks                       |
|---|---------------------------|--------|-------------------------------|
| Mainboard 68000<br>Stuecklisten                           | ME-11103720<br>11103720   | E<br>- | 2 pag.(Monoblock)<br>6 pages  |
| Mainboard 68000 BL<br>Stuecklisten                        | ME-11103721<br>11103721   | E<br>- | 2 pag.(Monoblock)<br>6 pages  |
| Mainboard 68000 (AB-S)<br>Stueckliste                     | ME11135172<br>ME-11135172 | A<br>- | (Traditional cell)<br>5 pages |
| Zellenprint (AB-S/ AB-L/ JB-C)<br>Stuecklisten            | ME-225615<br>225615       | D<br>- | (Monoblock)<br>2 pages        |
| Transducer board (PB-S/ PB-L/ GB-S/ JB-G)<br>Stuecklisten | ME-11103705<br>11103705   | B<br>- | (Monoblock)<br>2 pages        |
| Transducer board (AB-S)<br>Stueckliste                    | ME-238534<br>ME-238536    | F<br>- | (Traditional cell)<br>2 pages |

## 2.1.2 Essential characteristics

List of devices:

- Initial zero-setting;
- Zero-tracking;
- Combined semi-automatic zero-setting and subtractive tare balancing;
- Indication of unstable equilibrium;
- Determination stability of equilibrium;
- Automatic span adjustment with internal calibration mass; that operates:
 

|  |  |
|--|--|
| For AB-S/JB-C <ul style="list-style-type: none"> <li>- 1 hour after switch on and</li> <li>- 2 hours after switch on and</li> <li>- temperature: at least every 2 °C and</li> <li>- time: at least every 150 hours.</li> </ul> | For GB-S/PB-S/JB-G <ul style="list-style-type: none"> <li>- After switch on and</li> <li>- 2 hours after switch on and</li> <li>- temperature: at least every 2 °C and</li> <li>- time: at least every 150 hours.</li> </ul> |
|--|--|
- Semi-automatic span adjustment with internal calibration mass;
- Semi-automatic span adjustment with external calibration mass (only for class I);
- Auxiliary indicating with differentiated scale interval, (for AB-S models having a  $d < 0.1g$ , the auxiliary indicating is not differentiated);
- Piece counting;
- Percentage mode;
- Weighing unstable samples;
- Weight unit selection (kg/g/ct) or (g/mg/ct) (ct only for class I and II);
- Display indications other than primary indications;
- Acting upon significant faults;
- Checking the display;
- Underhook weighing.

## 2.1.3 Conditional parts

AC/DC adapter (ME-11106930 or ME-11132070):

- Input 110-240 V AC
- Output 12 V DC
- $I_{max}$  2.25 A
- $P_{max}$  27 W.

The interface section is located on the main board. The non-automatic weighing instrument may be equipped with one or more of the following protective interfaces that have not to be secured:

- RS232C;
- LC.

## 2.1.4 Non-essential parts

Display;  
Keyboard.

## 2.2 The mechanical assembly with weighing cell

### 2.2.1 Essential parts

| Description                             | Drawing number | Rev.     | Remarks                           |
|---|----------------|----------|-----------------------------------|
| Principle Schematic                     | SK-1461        | 14.05.98 | (Monoblock)                       |
| Principle Schematic                     | SK-220994      | 25.10.93 | (Traditional cell)                |
| Measuring cell (AB-S/ AB-L)             | ME-240866      | -        | Max = 310 g<br>(Monoblock)        |
| Measuring cell (GB-S /PB-S/ PB-L/ JB-G) | ME-240867      | -        | Max = 8100 g<br>(Monoblock)       |
| Measuring cell (AB-S/ PB-S/ PB-L/ JB-C) | ME-240866 A    | -        | Max = 610 g<br>(Monoblock)        |
| Messzelle (AB-S /AG)                    | ME-11505040 A  | -        | (Traditional cell)<br>Max = 220 g |

### 2.2.2 Essential shapes

See drawings:

- AB-S/ AB-L Balances, drawing number ME-240866;
- AB-S /AG Analysenwaagen, drawing number ME11505040 A;
- GB-S /PB-S/ PB-L/ JB-G Balances, drawing number ME-240867;
- AB-S/ PB-S/ PB-L/ JB-C Balances, drawing number ME-240866A.

## 3 Approval conditions



See chapter 1.3, essential shapes.

## 4 Seals and verification marks

See chapter 1.3, essential shapes.

## 5 CE-mark of conformity and inscriptions

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfils the requirements of article 1 of Annex IV.



|  |   |                                |                       |
|--|---|--------------------------------|-----------------------|
| Erstellt<br>Geändert                                   | 20.05.1998 H. Schaffner   | Hersteller-Code<br>Teileklasse | MacDraw™Pro Eichwesen |
| Mettler-Toledo GmbH<br>PO LabTec<br>CH-8606 Greifensee | Wir behalten uns alle Rechte an diesem Dokument und allen Beilagen vor. Der Empfänger anerkennt diese Rechte und wird die genannten Unterlagen nicht ohne unsere vorgängige schriftliche Ermächtigung Dritten zugänglich machen oder ausserhalb der Zwecks verwenden, zu dem sie ihm übergeben worden sind. |                                |                       |
| Hinweis AB-S and PB-S Balances                         |   |                                |                       |
| Position of verification- and securing stickers        |   | ME-240868                      |                       |
| Ersatz für Ø CD  | Ersetzt durch   | Blatt                          | METTLER-TOLEDO        |