



Steinbeis Transferzentrum Tübingen  
Arzneimittel – Kosmetika - Medizinprodukte

---

# Certificate

We hereby certify, that the below mentioned production lines meet

*all GMP-guidelines, EG –GMP-Annex 15 guideline  
and PIC/S guideline PI 006-3 requirements.*

Product lines: **D-, PUA, PTA45-, K-, M, PFA579,  
PFA579lift, PFA779lift,  
PBK989(APW) and PFK989(APW)**

Manufacturer: **Mettler-Toledo (Albstadt) GmbH  
Unter dem Malesfelsen 34  
72458 Albstadt  
Germany**

Tübingen, 12<sup>th</sup> January 2015

Prof. Dr. Ingrid Müller

Elke Weber, M.Sc. Dipl. Ing. (FH)

---

Steinbeis Transferzentrum  
Arzneimittel - Kosmetika - Medizinprodukte  
Leiter: Prof. Dr. Ingrid Müller  
Corrensstraße 34  
72076 Tübingen  
Tel.: 07071/37691  
E-Mail: [Ingrid.Mueller@stw.de](mailto:Ingrid.Mueller@stw.de), [Elke.Weber@stw.de](mailto:Elke.Weber@stw.de)



## 1. Description of the product line

### DN Line Bench scales

use of DMS cell (strain gauge)  
platform sizes [mm]: 240x300, 300x400, 400x500, 500x650, 600x800  
weighing capacity [kg]: 6, 15, 30, 60, 150, 300, 600

### DB-DCC

the DB-DCC-Line bench scales are built up hybrid – with a DMS-measuring cell and a lever arm system.

Platform size [mm]: 400x500, 600x800, 800x800, 1000x800  
Weighing capacity [kg]: 30, 60, 150, 300, 600

### DN- Line floor balances

Suitably for floor-mounted or pit-mounted installations

Use of 4 DMS measuring cells

platform size [mm]: 1250x1000, 1500x1250, 1500x1500, optional size  
weighing capacity [kg]: 300, 600, 1500, 3000

### PTA45

Pallet balances for weighing pallets simply

Use of 4 DMS measuring cells

platform size [mm]: 1260x600, optional size  
weighing capacity [kg]: 300, 600, 1500, 3000

### PUA579

extreme flat moveable balances

Use of 4 DMS measuring cells



Steinbeis Transferzentrum Tübingen  
Arzneimittel – Kosmetika - Medizinprodukte

---

platform size [mm]: 850x850, 1500x1250, optional size

weighing capacity [kg]: 300, 600, 1500

### **DRF/DSF Line**

flexible heavy scales, up to 3 modules,

suitable for floor-mounted and pit-mounted installations.

Use of 4 DMS measuring cells for each module

platform size [m]: 1,5 - 2 x 1,5 - 6

weighing capacity [kg]: 3000, 6000, 12000 (only DSF)

### **K-Line**

The K-Line contains bench- and floor scales.

Hybrid construction, electromagnetic strength compensation cell and lever arm system.

Calibrating capable resolution: up to 15000 / 32000e

Platform size [mm]: 280x350mm to 2000x1500mm

Weighing capacity [kg]: 15 kg to 6000kg

### **M-Line**

The M-Line contains bench- and floor scales.

Hybrid construction - digital DMS measuring cell and lever arm system.

Calibrating capable resolution: up to 3x3000e multi interval

Platform size [mm]: 280x350mm to 1500x1500mm

Weighing capacity [kg]: 15 kg to 3000kg

---

Steinbeis Transferzentrum  
Arzneimittel - Kosmetika - Medizinprodukte  
Leiter: Prof. Dr. Ingrid Müller  
Corrensstraße 34  
72076 Tübingen  
Tel.: 07071/37691  
E-Mail: [Ingrid.Mueller@stw.de](mailto:Ingrid.Mueller@stw.de), [Elke.Weber@stw.de](mailto:Elke.Weber@stw.de)



### **PFA579 line**

The line “PFA579” consists of floor scales, which can be upgraded and configured.

They can be installed as floor-mounted scales or as pit-mounted scales.

The four measuring cells (DMS measuring cells) are based on strain gauges.

The PFA579 line consists of stainless steel, type AISI 304/316.

Weighing range [kg]: 300-3000 kg

Platform sizes: between 700x400 and 2000x1500 mm

### **PFA579lift line**

PFA579lift floor scales are provided with a foldable plate load. They are suitable for routine cleaning. The weighbridges can be individually configured and so adapted according the requirements of the clients.



PFA579lift - opened load plate

Installation: Floor-mounted or pit-mounted

Technology: 4 DMS measuring cells

Material (Load plate and frame): Stainless steel AISI304- optional: AISI316

Weighing range: 300 – 3000kg

Platform sizes: Between 700x400 and 2000x1500 mm

Special sizes are available upon request.



### **PFa779lift line**

PFA779lift balances were developed in order to fulfill aspects of hygienic design. Therefore they can be used in areas where hygienic requirements

The weighing bridges are cleanable inside and outside because of the foldable plate load. The weighbridges can be individually configured and therefore adapted according to the requirements of the clients.



PFA779lift - opened load plate

Installation: Floor-mounted or pit-mounted

Technology: 4 DMS measuring cells

Material (Load plate and frame): Stainless steel AISI304- optional: AISI316

Weighing range: 300 – 3000kg

Platform sizes: Between 700x400 and 2000x1500 mm

Special sizes are available upon request.



### PBK9 series high precision bench scales

Product family: PBK989(APW) bench scales  
Technology: electromagnetic force compensation  
Material (load plate & frame): stainless steel AISI304  
Product option for load plate: stainless steel AISI316  
Surface quality load plate: Ra <math><1\mu\text{m}</math>  
Capacities: 0,6 – 300kg  
Dimensions: XS (210x250, load plate 130x160mm),  
A (280x350mm, load plate: 240x300mm),  
AB (280x350mm), B (400x500mm), CC (600x800mm)



### PFK9 series high precision floor scales

Product family: PFK989(APW) floor scales  
Technology: electromagnetic force compensation  
Material (load plate & frame): stainless steel AISI304  
Product option for load plate: stainless steel AISI316  
Surface quality load plate: standard: Ra <math><5\mu\text{m}</math>  
option: Ra <math><1\mu\text{m}</math>  
Capacities: 300 – 3000kg  
Dimensions: C (800x1000mm); D (1000x1250mm),  
E (1250x1500mm),  
ES (1500x1500mm)





## 2. General requirements for balances

Balances (weighing platforms) have to show the suitable measurement range and the required precision (EC-GMP guide<sup>1</sup>, chapter 3.40).

They have to be calibrated regularly, which has to be documented (EC-GMP guide, chapter 3.41).

The permitted tolerance must be provided for the respective weighing capacity, by consideration of the measuring inaccuracies, i.e. the still tolerated deviation of the debit value.

Working with raw materials, the equipment and the utensils used have to meet the requirements for surfaces in pharmaceutical production.

According to § 211.65 "Construction of the equipment"<sup>2</sup> of the FDA: *„(a) Equipment shall be constructed so that surfaces that contact components, in-process materials, or drug products shall not be reactive, additive, or absorptive so as to alter the safety, identity, strength, quality, or purity of the drug product beyond the official or other established requirements.“*

The cleaning ability is confirmed by the cleaning validation. Established cleaning instructions are necessary as a prerequisite for a cleaning validation.

The amount of permitted residuals whether active pharmaceutical ingredients or cleaning agents, is dependant on the preliminary manufactured product. These include the derivative product and the lot size of the derivative product. A first, general statement for not critical products can be made with the criteria "visual clean". According to the literature<sup>3</sup>, backlogs of 375 µg per 100 cm<sup>2</sup> are no longer visible.

---

<sup>1</sup> EC-GMP guide, chapter 3.40

<sup>2</sup> USA 21 CFR Part 211

<sup>3</sup> Buscalferrri et.al., Pharmind 62, Nr. 6 (2000)



### 3. Appraisal criteria for an optimal cleaning

#### *General*

In principle, those parts of equipment, which come into contact with the product, have to be cleaned well. Regarding the balances, these parts are the load plates.

Both the FDA inspection guideline for cleaning validation and the PIC guideline PIC/S 006-3 cite the visual criterion as one out of three possible acceptance criterions. From these the most appropriate criterion has to be chosen to appraise the cleaning success.

The weighing platforms of the product line D-, PUA, PTA45-, K-, M Line, PFA579, PBK989(APW), PFK989(APW), PFA579lift and PFA779lift were subjected to a qualified examination regarding "cleaning ability".<sup>1</sup> This means that the balances of the product lines D-, P- K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift must be free of visible residuals after cleaning. As a basis for this examination the works of Buscalferri, F., Assignment of the visibility limit of pharmaceutical active agents ("Bestimmung der Sichtbarkeitsgrenze von pharmazeutischen Wirkstoffen"), master thesis, Albstadt-Sigmaringen University, course of studies pharmaceutical technology (1999) and Fourman, G. L., Mullen, Determining Cleaning Validation Acceptance Limits for Pharmaceutical Manufacturing Operations, Pharm. Technol. 17 (4), 54 (1993) were used.

---

<sup>1</sup> The examinations and their evaluation were carried out by Prof. R. Ziegler.





### *Procedure*

A granulated material, which was coloured with Erythrosine, was used as a sample. The examinations<sup>1</sup> were carried out according to the “visually clean” criteria (please see “GMP Berater” 8.E.1.5). First the load plates of the product lines D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift were polluted, following by a cleaning step. A cleaning agent (P3-cosa PUR 80 Manufacturer: Ecolab GmbH & Co. OHG, Düsseldorf), which is commonly used in pharmaceutical production was taken and feigned according to different pollution degrees.

The cleaning success was then appraised visually.

In addition, a cast test was carried out before and after the cleaning to determine the complete microbial count.

The exact data can be taken from the SOP for cleaning and from the test report. These results show clearly, that the depletion degree meets in principle the hygienic requirements. In principle the depletion degree is dependent on the examined material and the specific requirements of the examination.

### *Results*

In regard to the cleaning, the load plates of the product lines, D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift, correspond to the visually clean criteria. The cast test for the determination of the complete microbial count showed a significantly lower complete microbial count after the cleaning.

<sup>1</sup> These experiments were performed by Prof. R. Ziegler



*Summary*

The platforms and terminals of the product lines, D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift are GMP-compliant. The cleaning of the parts, which come in contact with the product (load plate), has to be carried out well. There aren't any heavily accessible places in which dust could accumulate. The load plate is removable, so that the cleaning of the parts which are not in contact with the product is guaranteed.

The PFA579lift and the PFA779lift are easy to clean as well. Every part is easily accessible. The mechanic components are constructed according to the GMP guidelines.



## 4. Appraisal factors for qualification

### *General*

The PIC/S guideline PI 006-3 and the EG-GMP guide Annex 15 mention principles for qualification and validation.

Every machine or equipment which directly or indirectly influences the quality of the product shall be qualified. The machine or equipment shall be designed in agreement with the prevailing GMP guidelines. The machine shall be installed in agreement with the design specification and the functions shall be checked with the available documentation (functional qualification).

### *Procedure*

The available documentation of the METTLER TOLEDO product lines, D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift was checked to the effect of whether design qualification, installation qualification and functional qualification are feasible.



### *Results*

The documentation of the METTLER TOLEDO product lines, D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift is very detailed. Also an exact description of the balances with design drawings is available. The materials used are described precisely.

GMP-relevant documents are available (e.g. inspection certificate, CE-mark). Particulars regarding the maintenance are furnished.

### *Summary*

The documentation of the manufacturer of the METTLER TOLEDO product lines, D-, PUA, PTA45-, K-, M, PFA579, PBK989(APW), PFK989(APW), the PFA579lift and the PFA779lift is written in a very detailed way and offers the necessary conditions for the execution of qualification, as it is demanded by EG-GMP guide Annex 15 and PIC/S guideline PI 006-3.