SGS Portugal, S.A. Industrial Services Pólo Tecnológico de Lisboa R. Cupertino de Miranda, Lt 6, Piso 0 e 1 1600-546 Lisboa - Portugal T. (351) 21 710 42 00 F. (351) 21 715 75 26

Certificate of Conformity



acreditação

i0009 iSO/iEC 17020 inspeção

EU Final Assessment and Manufacturing Conformity

According to Directive 2014/68/EU

Certificate Nº: PTC21.02808,5308

Date:

2021.05.10

Name and Address of the Manufacturer: Mettler-Toledo GmbH

Im Hackacker 15, 8902 Urdorf

Switzerland

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 2014/68/EU. certifies that the Pressure Equipment design, construction and inspection, is in compliance with the requirements of the Pressure Equipment Directive stated, according module A2.

1155

Assessed According to Directive 2014/68/EU:

Module A2

Inspection Report Nº:

PTR21.03316.5308

Equipment ID:

Turbidity Sensor; DN50; SN: 1019902;

Model: InPro86X0ie

PS [bar]: 16.0

TS [°C]: -10 / 120

Category: I PED Data:

Fluid/Group: Gas - Gr.1

PT [bar]: 32 Date: 2021.04.27

Year of Manufacture: 2021

Valid Until: 2021.10.26

Carlos Felgueiras Technical Responsible



SGSPIP496/2 (2019-12-31)