

## **OIML Member State**

The Netherlands

## OIML Certificate N°R76/1992-NL1-03.16

Project number 303160 Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

**Issuing Authority** 

Name: NMi Certin B.V.

Address: Hugo de Grootplein 1, Dordrecht

Person responsible: P.P.M. van Enckevort

**Applicant** 

Name: Mettler-Toledo, Inc.

Address: 1150 Dearborn Drive

Worthington, Ohio 43085-6712 United States of America

Manufacturer of the certified type

Name: Mettler-Toledo, Inc.

Address: 1150 Dearborn Drive

Worthington, Ohio 43085-6712

**United States of America** 

Identification of the certified type

Type: PS15

 $n \le 3000$  divisions (per partial weighing range)

 $3 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$ 

e ≥ 1 g

maximum of 2 partial weighing ranges

Temperature range 0 °C / 40 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6274 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76** 

Edition 1992

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

Nederlands Meetinstituut Hugo de Grootplein 1 3314 EG Dordrecht Telephone +31 78 6332332

Telefax +31 78 6332309

NMi B.V.

(Chamber of Commerce no.27.228.701)

**Subsidiary companies:** 

NMi Van Swinden Laboratorium B.V. (27228703) NMi Certin B.V. (27.233.418) Verispect B.V. (27.228.700) This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



## OIML Certificate N°R76/1992-NL1-03.16

Project number 303160 Page 2 of 2

**OIML Member State** The Netherlands

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-03.16A that includes 53 pages. N° R76/1992-NL1-03.16B that includes 15 pages.

The Issuing Authority
P.P.M. van Enckevort
Manager Certification Delft

04 July 2003

The CIML Member G.J. Faber

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.